



महाराष्ट्र शासन

शासकीय दंतमहाविद्यालय व रुग्णालय, नागपूर

Government Dental College & Hospital, Nagpur

Govt. Medical College Campus, Medical Square, Nagpur - 440003, M.S. (INDIA)

Telephone No. 0712- 2743400 /2744496

Fax No. 0712- 2743400

Email - dean.gdcngp@gmail.com

GSTIN 27AAAGD1499M1Z1

जा.क्र.शादंमरुना/खरेदी विभाग/ Fellowship/ दरपत्रके / 3462 /2026

दिनांक :- 09/08/2026

Quotation Notice

To,

M/S.....

.....

.....

Subject :- Enquiry for Supply of Fellowship Dental Instrument/Equipments
Govt. Dental College & Hospital, Nagpur.

Ref's :- This Office Approved Notesheet Dated 25/08/2026

Dear Sir,

The undersigned invited sealed quotations for the items mentioned below/on/reverse or as per enclosed statement for the use of the Govt. Dental College & Hospital, Nagpur on the following terms & conditions.

१) The prices quoted should be free delivery to College & Hospital premises for local dealers and F.O.R. Nagpur for outsiders.

२) The Price quoted by inclusive of all taxes, duties if payable like customs, excise, CST, ST, and GST & inclusive transportation, installation, commissioning. The break-up of the taxes should also be shown separately where necessary. The sales tax & registration number should be quoted in your letter. Exemption of taxes, if on A.F. Forms etc.be separately stated.

३) The serial number of the items should not be changed while quoting rates. You may drop the item if not interested.

४) Rates should be quoted strictly for the item specified in the list and for standard quality of goods. In case of alternate offer, the detailed specifications, name of manufacturer or make etc must invariably be stated. Specifications other than specified in the schedule may be liable for rejection on even though lowest.

५) The quotation submitted will be valid for a period of six (६) months only from the date of acceptance.

8) The delivery of goods should be invariably be made within 2 weeks for indigenous items and 8 weeks for imported items from the date of confirmed order. **PENALTY ON LATE DELIVERY:** In the event of late delivery of goods, the purchaser will recover from contractor by way of penalty, a sum of equal to half percent of price of goods (The case of an order not exceeding Rs. 2.00 Lakh in Value subject to max. 90% and In case of an order of Rs. 2.00 Lakh and above subject to max. 4%) delivered late per week calculated from the next day after the agreed delivery period is over.

9) The quotation received after due date will not be accepted. The quotation should be submitted in sealed cover. Unsealed quotations which is not properly sealed will not be accepted. The last or receiving date of the quotation is 30/09/2024 upto 4.00 p.m.

10) Supply of stores should be made in one installment unless otherwise ordered place meal supply will not be accepted. Condition of advance payment through Bank, part payment etc. will not be accepted.

11) The price quoted is inclusive of all taxes, duties if payable like Customs/ Excise/CST/ST/GST the break up of the taxes should also be shown separately where necessary. The sales tax, GST and registration number should be quoted in your letter. Exemption of taxes on 'AF' / Octroi exemption form etc be separately stated.

12) Quotations if asked with samples if not accompanies with sample will be liable for rejection even they are lowest. Samples should be sent with a labels attached quoting our ref.no of enquiry and item no etc.

13) You are also required to submit the security deposit at the rate of 3 % of the cost of equipment in the form of Bank Guarantee for two years (the name of Dean, Govt. Dental College & Hospital, Nagpur) and 5 % performance Bank Guarantee for 8 year (after warranty period) for equipment above one lakh only.

14) The tenderer will have to agree to enter into annual maintenance contract (AMC) @ of 9% of the ordered value of the equipment. Where required tenderer will have to agree for comprehensive maintenance contract (CMC) inclusive of all spares @4% of the ordered value of the equipment. The period of such AMC/CMC will be of 2 years after completion of warranty period. In case of non-compliance of AMC/CMC the supplier will be liable to pay a penalty. Such penalty shall be recovered from the amount of bank guarantee.

15) Comprehensive warranty (Including supply of spares) All inclusive warranty period shall be of 2 years from the date of commissioning of all equipment supplied as certified by the consignee. The successful bidder must ensure 900% uptime during warranty period. In case of down time, warranty period will be extended for period of down time.

१४) Initial payment of ९०% of the contract value will be made on delivery, installation and submission of security deposit of ३% (BG) & १०% of the contract value will be paid on satisfactory commissioning of the machinery within ३० days.

१५) The tenderer should install the equipment in the concerned department of the institution, give demonstration free of cost and train the staff of the department (no. as specified by user)

१६) The Decision of the Dean, Govt. Dental College & Hospital, Nagpur will be final and binding. The Dean, Govt. Dental College & Hospital, Nagpur reserves right to reject any or all Quotations without assigning any reason.

१७) Very Important :- In case you are not interested in quoting your rates; in reply to this enquiry a line in reply is must; failing are liable for delisting their names for further enquiries from our list and no further request in this matter will be entertained.

१८) Literature, Instruction manual, working leaflet etc may also be sent with quotation.

१९) Authorization certificate, Brochure, Catalogue, User List may also submit along with quotation. The Photostat copies of the firm registration should be attached with the quotation. Authorization letter: A dealer quoting on behalf of manufacturer must attach authorization letter from the manufacturer to sell the equipment and to do after sales services. Sub dealer certificate will not be accepted.

२०) Technical Specifications – Compliance Chart with variation/deviation.

IMP:- General Terms & Conditions;

a) Bidder should fill the **Technical Bid & Price Bid** in separate envelop.

b) Envelope (Technical & Price bid) consists of authorization letter & compliance chart are mandatory. Firm registration certificate, GST Registration, Product catalogue, ISO or ISI certificate. & price of equipment/machinery

c) Price of equipment must be quoted inclusive of all taxes.

d) A hands on working demonstration is essential before opening envelope

e) Unsatisfactory performance of demonstration will disqualify the tender.

Details as below;

Purchase of Equipments for the year २०२६-२७.

Department :- Oral Surgery, Prostodontics

Sr.No.	Name of Machinery & Equipment	Department	Qty	Total Rate (including all Taxes)
1	Minor Surgical Set for disimpaction	Oral Surgery	1	
2	Exfraction Forceps set		1	
3	Root Elevator Set		1	
4	Luxators Set		1	
5	Durr Compressor For CAD CAM Machine	Prostodontics	1	

NOTE:- Specifications of above equipments are attach.



Dean,
Govt. Dental College & Hospital,
Nagpur.

Specifications

9. Minor Surgical Set for disimpaction Specifications

Specifications :- Specification Of Minor Surgical Set For Disimpaction

Diagnostic Instruments

Sr. No.	Surgical Set (one unit) Includes	Quantity
1	Mouth mirror	10 Each
2	Periodontal probe / Straight probe	10 Each
3	Tweezer	10 Each

Surgical Instruments

Sr. No.	Surgical Set (one unit) Includes	Quantity
1	Moons probe	10 Each
2	BP Handle No. 3	10 Each
3	Periosteal elevator	10 Each
4	Austin tissue retractor – Regular	10 Each
5	Kocher Langenbeck retractor 25 cm	10 Each
6	Minnesota retractor	10 Each
7	Root tip elevators	10 Each
8	Lucas curette	10 Each
9	Mouth props	10 Each
10	Needle holder 6"	10 Each
11	Scissors - northbent - Suture cutting	10 Each
12	Bone file	10 Each
13	Bone rongeur	10 Each
14	Suction cannula - Fraser No. 3	10 Each
15	Dissecting forceps - Adson 7" tooth	10 Each
16	Towel clip	10 Each
17	Curved artery forceps×2	10 Each
18	Straight hand piece	10 Each
19	Bone cutting burs (Carbide HP 701, 702, 703, Round burs)	10 Each
20	Aluminium General box – Anodized small	10 Each

2. Extraction Forceps set Specifications

Specification of Extraction Forcep Set

Sr. No.	Name of Extraction Forcep Set	Quantity
Maxillary		
1	Anterior forcep	10 Each
2	Premolar forcep	10 Each
3	Molar forcep (right & left)	10 Each
4	Cowhorn forcep (right & left)	10 Each
5	Third molar forcep	10 Each
6	Root forcep	10 Each
7	Bayonet forcep	10 Each
Mandibular		
1	Anterior forcep	10 Each
2	Premolar forcep	10 Each
3	Molar forcep	10 Each
4	Cowhorn forcep	10 Each
5	Third molar forcep	10 Each

3. Root Elevator Set Specifications

Root Elevator Specifications

Material: High-quality surgical stainless steel (corrosion-resistant and autoclavable)

Components:

Handle

Shank

Working blade (tip)

Handle: Ergonomic, non-slip grip
Usually octagonal or round with serrations
Provides good control and torque
Shank: Strong and rigid
Straight or angled depending on the design
Transfers force efficiently from the handle to the blade
Blade (Working End):
Thin, sharp, and tapered
Concave on one side and convex on the other (for straight elevators)
Available in various widths (approximately 2–5 mm)
Designed to fit between the tooth/root and alveolar bone
Sterilization: Steam autoclave compatible (121–134°C)
Common Types and Sizes
Straight elevator (e.g., No. 301, 34S)
Coupland elevators (1, 2, 3)
Cryer (East-West) elevators (Left and Right)
Warwick James elevators (Left, Right, Straight)
Apexo elevators (small, medium, large)
Ideal Requirements
Strong enough to withstand controlled force without bending.
Sharp enough to engage the root securely.
Well-balanced and comfortable to hold.
Easy to sterilize and resistant to corrosion.

8. Luxators Set Specifications

Specification of Veecare Angled Luxators

1. Material: Surgical-grade French stainless steel
2. Tip Configuration: Angled working end
3. Available Sizes: 2mm, 3mm, 4mm
4. Handle Design: Anatomically contoured ergonomic grip
5. Surface Treatment: Passivated for corrosion resistance
6. Sterilization: Fully autoclavable
7. Clinical Use: Atraumatic tooth extraction and PDL severance
8. Brand: Veecare

5. Specifications Of Durr Compressor For CAD CAM Machine

COMPRESSOR FOR CAD CAM MACHINE

The compressor should be a 1HP- 2HP medical grade compressor with radiator, cooling fan, dryer and silicone column filter.

The compressor should be noiseless and provide oil free and dry compressed air.

The compressor should have a pressure range of 6-7,8 bars.

The compressor should have a noise level of approx. 70 dB(A) at 5 bars.

The Compressor should work at a power supply of 230 V, 50-60 Hz

The compressor should have an output of 215/240 L/min at 5 bar.

The compressor should have 4 cylinders.

The compressor must have autohead air release valve and automatic cut off.

The compressor should have a tank capacity of atleast 50 L.

The compressor should have a membrane dryer to prevent moist environments.

The inner surface of the compressor tank should be coated with epoxy to prevent rusting.

The compressor should provide hygienic, tasteless and odourless compressed air.